

Kinetigram Customer Release Notes

1 Administrative Information

Software Version: V1.1.0

Manufacturer: Bioniks
2156 Central Avenue
Alameda, CA 94501 USA
help@bioniks.net
(510) 747-1947
<http://www.bioniks.net>

Document #: 320-00201-DO-03 Rev 3 07/16

2 Indications for Use

The Kinetigram is Software as a Medical Device intended to be used with the Microsoft Kinect for Xbox One to evaluate upper extremity joint function using 3D kinematic data capture in adults. The product is intended for use by medical professionals amongst other available patient parameters as a motion analysis tool.

CAUTION (United States): Federal law restricts this device to sale by or on the order of a physician.

3 Issues Fixed

N/A – No fixed issues from the previous release version.

4 Known Issues

N/A – There are no known issues that impact the use of Kinetigram.

5 Cautions

All medical devices, including Kinetigram has inherent potential risk associated with its use. However, it has been determined, based on the likelihood and severity of those risks that the benefit if its intended use outweigh the potential risk.

EU/International: Exclusively for Investigational Use.

Please see the Instructions for Use document for a full listing of the cautions.